

Country-Specific Glucose Monitor List

Country Name: *Ireland*

Important Information

This is a non-comprehensive list, current as of August 2025. Absence of a specific glucose monitor or test strips from this list does NOT imply compatibility or incompatibility with EXTRANEAL (Icodextrin) PD solution. Similarly, other glucose-measuring technologies which are not listed below (such as continuous glucose monitoring systems) may or may not be compatible with EXTRANEAL. **Always contact the device manufacturer for current information.** If the manufacturer cannot provide information regarding compatibility of the device with Icodextrin and Maltose, Vantive does NOT recommend that EXTRANEAL patients use the product. Vantive reserves the right to change this list without notice and does not represent that it includes all potentially incompatible products.

The glucose monitors manufacturers listed have certified to Vantive that they have tested (as per ISO 15197) their monitors with Maltose and Icodextrin to Vantive's recommended limits: 278 mg/dL (Maltose) and 1094 mg/dL (Icodextrin). The manufacturers certified that their "green" monitors mentioned below showed no interference of blood glucose readings. Please note that the compatibility list below is only for brand-name monitors used with the corresponding brand-name test strip. If a brand-name monitor is used with another manufacture's test strip, **always contact the test strip manufacturer for current information.**

This list is compiled from a search via: Internet, Literature, Vantive internal studies, information from government agencies, test strip leaflets, safety alerts, and direct information from the product manufacturers. While efforts have been made to provide accurate and current information, Vantive does not manufacture these glucose monitors or test strips and thus does not guarantee the initial or continued accuracy of this information. Please contact the manufacturer(s) of the glucose monitor and test strip to obtain the latest compatibility information before using in conjunction with EXTRANEAL (Icodextrin) Peritoneal Dialysis (PD) solution.

1. EXTRANEAL (Icodextrin) PD solution contains Icodextrin. Maltose, a metabolite of Icodextrin, may interfere with certain glucose monitors or test strips. This interference will result in a falsely elevated glucose reading using these monitors or test strips.
2. This interference may mask true hypoglycemia or lead to the erroneous diagnosis of hyperglycemia. Thus, a blood glucose reading within or above the normal range in a patient on EXTRANEAL (Icodextrin) PD solution, using these monitors or test strips, may mask true low blood sugar. This may cause a patient or health care professional to not take the appropriate steps to bring the blood sugar into a normal range. Alternatively, a falsely elevated blood glucose reading could cause a patient to get more insulin than needed. Both of these situations can lead to life-threatening events, including loss of consciousness, coma, neurological damage or death.
3. The dialysis unit or patient should contact the manufacturer of the glucose monitor and/or test strips to determine if the monitor or test strips they are using are subject to interference by Icodextrin or Maltose. Also, consult the product information included with the glucose monitor and test strips.
4. The following list is for reference only. This list does not imply recommendation of these glucose monitors or test strips.
5. Identified compatibilities are shown in the table below. ONLY glucose-specific monitors and test strips should be used with patients on EXTRANEAL (Icodextrin) PD solution. Contact the manufacturer to verify that the test strips and monitor are glucose specific. The list provides the contact information of the more common, major brand manufacturers.

COUNTRY-SPECIFIC GLUCOSE MONITORS

Updated Ireland August 2025

Glucose Monitor Brand	Compatible with Extraneal (Icodextrin) PD solution (Glucose-specific)	Test Type*	Manufacturer
FreeStyle Freedom	Yes	GDH-FAD	Abbott Diabetes Care www.abbottdiabetescare.com https://freestylediabetes.ie Phone: +1800 77 66 33
FreeStyle Freedom Lite	Yes	GDH-FAD	
FreeStyle Lite	Yes	GDH-FAD	
FreeStyle Libre1	Not Recommended	GO	
FreeStyle Libre Pro1	Not Recommended	GO	
FreeStyle Libre 21	Not Recommended	GO	
FreeStyle Libre 31	Not Recommended	GO	
FreeStyle Optium Neo	Yes	GDH-NAD	
FreeStyle Optium Neo H	Yes	GDH-NAD	
FreeStyle Papillon Vision	Yes	GDH-FAD	
FreeStyle Precision Neo	Yes	GDH-NAD	
FreeStyle Precision Pro	Yes	GDH-NAD	
Optium Xido Neo	Yes	GDH-NAD	
Precision Xceed Pro	Yes	GDH-NAD	
Assure Platinum	Yes	GO	Arkray, Inc. www.arkray.co.jp www.arkray.eu Phone N/A
GLUCOCARD 01	Yes	GO	
GLUCOCARD 01-mini	Yes	GO	
GLUCOCARD 01-mini plus	Yes	GO	
GLUCOCARD MX	Yes	GDH-FAD	
GLUCOCARD S	Yes	GDH-FAD	
GLUCOCARD SM	Yes	GDH-FAD	
GLUCOCARD Vital	Yes	GO	
GLUCOCARD X-meter	Yes	GDH-FAD	
GLUCOCARD X-mini	Yes	GDH-FAD	
GLUCOCARD X-mini plus	Yes	GDH-FAD	
GLUCOCARD Σ	Yes	GO	
GLUCOCARD Σ-mini	Yes	GO	
ReliOn Confirm	Yes	GO	
ReliOn micro	Yes	GO	
ReliOn Prime	Yes	GO	
GLUCOCARD S onyx	Yes	GDH-FAD	
PocketChem BG	Yes	GDH-FAD	
GLUCOCARD W	Yes	GO	
GLUCOCARD PlusCare	Yes	GDH-FAD	
GLUCOCARD Prime	Yes	GDH-FAD	
GlucoCard G+	Yes	GO	
GlucoCard G+ Link	Yes	GO	
GlucoCard Σ+	Yes	GO	
GlucoCard Σ+ Link	Yes	GO	

Glucose Monitor Brand	Compatible with Extraneal (Icodextrin) PD solution (Glucose-specific)	Test Type*	Manufacturer
Contour Contour Next Contour Next EZ Contour Next One Contour Plus Contour Plus One Contour TS Contour XT Contour Next (Connected) Contour Care Contour Plus Blue Contour Plus Elite Contour Next GEN	Yes Yes Yes Yes Yes Yes Yes Yes Yes Yes Yes Yes Yes Yes	GDH-FAD GDH-FAD GDH-FAD GDH-FAD GDH-FAD GDH-FAD GDH-FAD GDH-FAD GDH-FAD GDH-FAD GDH-FAD GDH-FAD GDH-FAD GDH-FAD	Ascensia Diabetes Care (formerly Bayer Healthcare) www.ascensia.com https://diabetes.ascensia.ie Phone: +1890 920 111
OneTouch SelectPlus OneTouch SelectPlus Flex OneTouch SelectPlus Simple OneTouch Select Simple OneTouch Ultra 2 OneTouch UltraEasy OneTouch UltraMini OneTouch UltraPlus Flex OneTouch UltraPlus Reflect OneTouch UltraVue OneTouch Verio OneTouch VerioFlex OneTouch VerioIQ OneTouch VerioPro+ OneTouch VerioVue OneTouch Verio Reflect	Yes Yes Yes Yes Yes Yes Yes Yes Yes Yes Yes Yes Yes Yes Yes Yes Yes	GO GO GO GO GO GO GO GO GDH-FAD GDH-FAD GO GDH-FAD GDH-FAD GDH-FAD GDH-FAD GDH-FAD GDH-FAD	Lifescan, Inc. www.lifescan.com www.myonetouch.co.uk Phone: 0800 121 200
Nova Max Plus Nova Max Link StatStrip Hospital Glucose Meter StatStrip Hospital Glucose and Ketone Monitoring System StatStrip Xpress Glucose Meter StatStrip Xpress Glucose and Ketone Monitoring System StatStrip Xpress2 Glucose and Ketone Monitoring System Nova Pro Glucose/Ketone Meter StatStrip Glucose Hospital Meter System V2.0 StatStrip Glucose/Ketone Hospital Meter System V2.0	Yes Yes Yes Yes Yes Yes Yes Yes Yes Yes	GO GO GO GO GO GO GO GDH-FAD GO GO	Nova Biomedical www.novabiomedical.com uk-info@novabio.com Phone: +44 1928 704 040
Accu-Chek Active Accu-Chek Aviva Accu-Chek Aviva Combo	Yes Yes Yes	Mut Q-GDH Mut Q-GDH Mut Q-GDH	Roche Diagnostics https://www.accu-chek.co.uk/

Glucose Monitor Brand	Compatible with Extraneal (Icodextrin) PD solution (Glucose-specific)	Test Type*	Manufacturer
Accu-Chek Aviva Connect Accu-Chek Aviva Expert Accu-Chek Aviva Insight Accu-Chek Aviva Plus ² Accu-Chek Aviva Solo Accu-Chek Guide Accu-Chek Guide Me Accu-Chek Inform II Accu-Chek Instant Accu-Chek Instant S Accu-Chek Nano ² Accu-Chek Nano SmartView ² Accu-Chek Mobile Accu-Chek Performa Accu-Chek Performa Combo Accu-Chek Performa Connect Accu-Chek Performa Insight Accu-Chek Performa Solo Accu-Chek Guide Solo Cobas Pulse	Yes Yes Yes Yes Yes Yes Yes Yes Yes Yes Yes Yes Yes Yes Yes Yes Yes Yes Yes	Mut Q-GDH Mut Q-GDH Mut Q-GDH Mut Q-GDH Mut Q-GDH GDH-FAD GDH-FAD Mut Q-GDH GDH-FAD GDH-FAD Mut Q-GDH Mut Q-GDH Mut Q-GDH Mut Q-GDH Mut Q-GDH Mut Q-GDH GDH-FAD GDH-FAD	https://www.accu-chek.co.uk/?diabetescare=IE Phone: 1 800 709 600 or 01 5410379
TRUE METRIX TRUE METRIX GO TRUE METRIX Air TRUEresult ³ TRUEresult Twist 2 ³ TRUEyou ³	Yes Yes Yes Not Recommended Not Recommended Not Recommended	GDH-FAD GDH-FAD GDH-FAD GDH-FAD GDH-FAD GDH-FAD	Trividia Health www.trividiahealth.com www.trividiahealth.co.uk Phone: 0800 689 5035
Rightest Blood Glucose Monitoring System GM700 Rightest Blood Glucose Monitoring System GM700 Pro Rightest Blood Glucose Monitoring System GM700S Rightest Blood Glucose Monitoring System GM700SB Rightest Blood Glucose Monitoring System GM700 Pro 2 Rightest Blood Glucose Monitoring System GM720 Rightest Blood Glucose Monitoring System GM722 Rightest Blood Glucose Monitoring System Max	Yes Yes Yes Yes Yes Yes Yes Yes Yes Yes	GDH-FAD GDH-FAD GDH-FAD GDH-FAD GDH-FAD GDH-FAD GDH-FAD GDH-FAD GDH-FAD	Bionime Corporation www.bionime.com Phone: +886 4 2369 2388 European Office www.bionime.ch Phone: +41 71 722 9840

Glucose Monitor Brand	Compatible with Extraneal (Icodextrin) PD solution (Glucose-specific)	Test Type*	Manufacturer
Rightest Blood Glucose Monitoring System Max Plus	Yes	GO	
Rightest Blood Glucose Monitoring System Wiz	Yes	GO	
Rightest Blood Glucose Monitoring System Wiz Plus	Yes	GO	
Rightest Blood Glucose Monitoring System GM550	Yes	GO	
Rightest Blood Glucose Monitoring System GM280	Yes	GO	
Rightest Blood Glucose Monitoring System GM280B	Yes	GO	
Rightest Blood Glucose Monitoring System GT333	Yes	GO	
Rightest GT333 Blood Glucose Monitoring System	Yes	GO	
Rightest Blood Glucose Monitoring System GM300	Yes	GO	
Rightest Blood Glucose Monitoring System GM250	Yes	GO	
Rightest Blood Glucose Monitoring System GM260	Yes	GO	
GE100 Blood Glucose Monitoring System			
Diathrive+ Blood Glucose Monitoring System	Yes	GO	
Diathrive+ Bluetooth Blood Glucose Monitoring System	Yes	GO	
iGlucose Blood Glucose Monitoring System	Yes	GO	
GE333 Blood Glucose Monitoring System	Yes	GO	
Quintet AC Blood Glucose Monitoring System	Yes	GO	
Mylife Pura Blood Glucose Monitoring System	Yes	GO	
Mylife Unio Blood Glucose Monitoring System	Yes	GDH-FAD	
Mylife Unio Neva Blood Glucose Monitoring System	Yes	GDH-FAD	
Mylife Unio Cara Blood Glucose Monitoring System	Yes	GDH-FAD	
	Yes	GO	

Glucose Monitor Brand	Compatible with Extraneal (Icodextrin) PD solution (Glucose-specific)	Test Type*	Manufacturer
LineaD SIRIO Blood Glucose Monitoring System	Yes	GO	
LineaD EVO Blood Glucose Monitoring System	Yes	GDH-FAD	
LineaD ORO HR Blood Glucose Monitoring System	Yes	GDH-FAD	
LineaD ORO Blood Glucose Monitoring System	Yes	GDH-FAD	
LineaD ORO24 Blood Glucose Monitoring System	Yes	GDH-FAD	
Rightest Blood Glucose Monitoring System Max Tel	Yes	GDH-FAD	
Rightest Blood Glucose and Ketone Monitoring System Max Dual			
GlucoRx Nexus	Yes	GDH-FAD	GlucoRx www.glucorx.ie Phone: 00353 1 6950401
GlucoRx Nexus Blue	Yes	GDH-FAD	
GlucoRx Nexus Pearl	Yes	GDH-FAD	
GlucoRx Nexus Mini Ultra	Yes	GDH-FAD	
GlucoRx Nexus Voice	Yes	GDH-FAD	
GlucoRx GO	Yes	GDH-FAD	
GlucoRx GO Professional	Yes	GDH-FAD	
GlucoRx HCT	Yes	GDH-FAD	
GlucoRx Smart	Yes	GDH-FAD	
GlucoRx Vivid	Yes	GDH-FAD	
GlucoRx Vivid Blue	Yes	GDH-FAD	
GlucoRx Q	Yes	GO	
GlucoRx Q Blue	Yes	GO	
GlucoRx X6	Yes	GDH-FAD	

¹ These Abbott monitors (together with the Dexcom G4 Platinum, G5 Mobile, G6, G7 and the Medtronic iPro2) are the leading Continuous Glucose Monitor Systems (CGM) and are **not recommended** by their manufacturers for use in dialysis patients, and despite having a GO based chemistry currently there is not definitive scientific evidence for their Extraneal compatibility. Consult manufacturer for additional information.

² The ACCU-CHEK Nano (not Aviva or Performa) and ACCU-CHEK Aviva Plus glucose systems are available within the United States ONLY and use test strips branded as ACCU-CHEK SmartView and ACCU-CHEK Aviva Plus, respectively. These systems use the Mut Q-GDH (**Compatible**) strips. Consult manufacturer for additional information.

³ TRUEresult (GDH-FAD), TRUEresult Twist 2, TRUEyou, Truetrack monitors were not tested with the Vantive recommended limits for Icodextrin and Maltose. So, these monitors are **not recommended** to use for dialysis patients.

Test Type*

GO = Glucose oxidase

GDH-NAD = Glucose dehydrogenase with nicotinamide-adenine dinucleotide

GDH-FAD = Glucose dehydrogenase with flavin-adenine dinucleotide

Mut Q-GDH = Glucose dehydrogenase with pyrroloquinolinequinone modified to eliminate maltose interference

Please see full prescribing information for EXTRANEAL (Icodextrin) PD solution.

Vantive and EXTRANEAL are trademarks of Vantive Health LLC.
Other trademarks or products appearing herein are the property of their respective owners.